Office of the Principal NRS Medical College, Kolkata

Memo No: NMC/Store/ 1034

Date: 28/02/2019

<u>Tender Notice</u> (NIT NO. 76 /2018-19)

Sealed Tender is invited by the Principal, N. R. S. Medical College, Kolkata – 14 for supply of 1) Sperm Counting Chamber 2) Ph Meter 3) Anti vibration Table 4) Test Tube Warmer digital for the Department of G & O for IVF Centre of this Institution. Tender form and details/specification can be obtained from the Departmental website www.wbhealth.gov.in from 01/03/2019. Last date of submission of Tender is 11/03/2019 up to 1 P.M. Tender will be opened on same date at 2.30 P.M in the office of the Principal, N. R. S. Medical College. Details are available in the Department of Health & Family Welfare, Govt. of West Bengal official Website: www.wbhealth.gov.in & College website (www.nrsmc.edu.in)

Principal
NRS Medical College,
Kolkata-14

Dated: 28/02 2019

Memo No: NMC/Store/ 1034 1 (4)

- 1. The Medical Superintendent-cum-Vice Principal, NRSMCH, Kolkata
- 2. Accounts Officer, NRS Medical College, Kolkata
- 3. I. T. Cell, Swasthya Bhawan, Salt-lake, for publication in the website.
- 4. Notice Board.

Principal
NRS Medical College
Kolkata-14

Terms and Condition

- 1. Only Manufacturers/direct Importers/Authorized Distributors are eligible for quoting.
- should be quoted in both figure and word strictly as per provided accounting unit specification along with sample/mentioned liveable charges inclusive of all should be All rates quoted department, packing the concerned delivery charges to door (including charges, installation charges), and all other foreseeable incidental charges are exclusive of all applicable taxes should be mentioned separately.
- 3. The Tendering authority reserves the right to withdraw any item from the Tender at any stage. In such Contingency the selection of such item, if already made in favour of any Tendered shall be treated as cancelled. The Tendering authority also reserves the right to accept / reject / split any.
- 4. The Tenderer shall be bound for total replacement of the supplied quantity with a fresh supply of equivalent quantity if the item(s) are found to be defective/not as per given specification.
- 5. Payment will be made only after the delivery of the articles and satisfactory certification by an authority fixed by the administration. No request for advance or part payment will be considered.
- 6. In case of rejection of any supply on quality grounds or failure or refusal to replace the contracted items within the specified time or according to specifications/quality, the College/Hospital administration will have the authority to recover the losses, if any, by making deductions from any pending claim of the supplier in respect of same tender or any past/future tender.
- 7. The Specification and Brochure is to be submitted along with the Tender.
 - 8. The Tenderers must accompany all the papers/documents along with the Tender.
- 9. The Tenderers must submit their tender in two bid method. Main envelope should contain Technical bid in Envelop A and Financial bid in Envelop B. all the envelops must be properly sealed before submission.
 - 10. Price bid of those bidders will only be opened who will qualify in technical scrutiny by tender selection committee.
- 11. Earnest money amounting to Rs. 5,000/- (Five thousand) only must be deposited in the form of Demand Draft from any Nationalized Bank in favour of 'Principal, N. R. S. Medical College' payable at Kolkata by L1 bidder before collection order.
- 12. The performance bank guarantee will be mandatory for all suppliers and will not be waived in any case. The successful renderers shall be required to furnish the 'performance bank guarantee' from any Nationalized/scheduled Bank acceptable to Govt. Of West Bengal for one year @ 10% of quoted base rate per item for which the tenderer has been selected as supplier.
- 13. The tendering authority has the liberty to ask for **performance demonstration** of any item during any stage of the tender process, before placing order for supply. In case of dissatisfaction, the authority will have the right to disapprove the offer of L1 bidder and go for the next bidder. Orders for supply of the approved products will be placed with the successful tenderers after execution of agreements and such supply shall have to be made in such installments as may be fixed or spread over the period to be specified in the supply orders to be made in pursuance of the agreements. The equipment is to be transported in such packaging so as to avoid damage to primary package of manufacturer. The successful tenderer will have to supply within the specified time schedule that had been assured at the time of selection as supplier.

14. Guarantee / Warranty Period:

The machine/Instrument should be covered by free on-site comprehensive Warranty for one year after successful installation. The tenderers must quote for 5 (Five) years' comprehensive on-site AMC of entire system (including all spares & labour) after the date of completion of free warranty period.

15. Performance Bank Guarantee:

The performance bank guarantee will be mandatory for all suppliers and will not be waived in any case. The successful tenderers shall be required to furnish the 'Performance Bank Guarantee' from any Nationalized/Scheduled Bank acceptable to Govt of West Bengal for One year @ 10% of quoted base rate per item for which the tenderer has been selected as supplier. The performance bank guarantee will have to be renewed after completion of Free Warranty period, for additional 5 (Five) years under the paid CAMC period before transfer of the CAMC amount to bidder's account.

16. In case of more than one L 1 bidder, weightage will be given to document number 7 of the following list.

List of essential documents to be submitted

- 1. Application submitted in Annexure-I
- 2. Copy of GST Registration certificate showing Registration No. (not provisional)
- 3. Copy of Valid Trade license
- 4. Copy of PAN Card of the bidder
- 5. Copies of Income Tax return for the last three years
- 6. Copies of Profit & Loss A/c and Balance Sheet for the last three years .
- 7. Credential of the bidder in successful supply to Govt. hospitals

Principal

NRS Medical College

Kol-14

ANNEXURE-I

APPLICATION FORMAT

To The Principal, Nil Ratan Sircar Medical	letter pad with full address and contact no etc) College, Kolkata-700014.	
Sub: Tender No(Nan	me of the Item -1)	-2)
	ipment for department of, Nil Ratan Sircar	
submit all the necessary information and		
1. That the application is made by me / us In the capacity ofauthorization letter from the Company is	duly authorized to submit the offer. The	
declare that we shall abide by it for through. We are offering rate for the following it	laid down in the NIT mentioned above and ghout the tender period and its extensions, if any. tem /items and assured supply to the Health & t of West Bengal as per Table-1 (Instrument	
	having its office at	
	(Address, contact no and e mail address) as Others (strike out whichever is not applicable) as	
	the distributor/C&F/Liasoner& other documents	
5. In the event of being selected, supply wexcepting the condition which is beyond of 6. We understand that:		
(a) Tender Selection Committee/ H&FW Dunder this project.	Dept can amend the scope & value of the contract bid	
(b) Tender Selection Committee/ H&FW E without assigning any reason. Date:-	Dept reserves the right to reject any application	
Contact details -	Signature of applicant including title and capacity in which application is made.	

Mobile: E mail ID: Product Name: Sperm Counting Chamber

Advantages:

- Applied spermatozoa are uniformly distributed and monolayered, and are observed in one focal plane.
- Dilution is unnecessary even with concentrated specimens. Analysis is done directly from original specimen in its natural environment.
- All spermatozoa acquire friction free, horizontal movement and are always examined under constant conditions.
- The specimen can be analyzed quickly as an office procedure while the patient is waiting, and even by an
 inexperienced person.
- Accuracy of analysis is enhanced through the elimination of the various steps required by the usual hemocytometric technique. In addition, the fact that sperm motility is examined each time under identical conditions further increases accuracy. Errors incurred by uncontrolled pressure applied to the cover slip are thus avoided.
- The 10 micron depth of the Sperm Chamber is ideal for still or movie camera photomicrography, as it approximately matches the field depth of the objective used in semen analysis.
- The Chamber is quickly and easily available for reuse. In a busy laboratory a large number of tests per hour can be made by a single technician with minimal technical and material requirement

Professor & Head

Professor & Head

Professor & Gynae. & Obts

Department of Gynae. & NRSMCH. KOL.

Lauren duracija.

TECHNICAL SPECIFICATION OF pH METER 01 No.

1. General Features:

- 1. pH Meter readings should have excellent calibration response and repeatability.
- 2. pH measuring range: 0.0 0-14.0 Ph.
- 3. pH resolution range: 0.01-0.05.
- pH accuracy: ±0.05.
- 5. Continuous ph monitoring (reading and recording every 30 minutes).
- 6. Track up to 5-10 trackpods. Real time ph monitoring and east data access.

2. Calibration:

- 1. Calibration: Digital automatic calibration and automatic off function.
- 3. Increased data storage capacity.
 - 1. Data Export: Data export to PC with dedicated software.

46

Professor & Head

Dr. Chandana Das

Professor Gunae & Obts

Cobts

Depar

Sensain Museugir.

Product Name: Anti-vibration table

- Anti-vibration table is for ICSI procedures.
- It is designed to meet the requirements for all available models of inverted microscopes.
- The surface of the table comes in Hi-MACS or stainless steel
- It is easy to clean, hygienic and does not absorb liquids or odours.
- Standard size w x d x h: 1150 mm x 720 mm x 790-830 mm

Laming Mukuya 21/2/19

Professor & Head

Professor & Head

Dr. Chandana Das

Dr. Chandana Das

Obts

Department of Gynae. & Obts

NESMCH. KOL.

TECHNICAL SPECIFICATION DIGITAL TEST TUBE WARMER.

Should have visible digital LED display indicates the temperature level in either

- 1. Celcius or Fahrenheit.
- 2. Temperature precise and reliable within the incubator's range of ambient to 50 °C.
- 3. The front panel continuous observation of the test tubes' contents.
- 4. The heater maintain an operating temperature of 37 °C.
- The light on the front of the heater should indicates that the unit is powered. When the light flashes, the heater is at its operating temperature.
- 6. Temperature level should adjusted directly on the display in increments of 0.1°C.
- Designed to offer minimal resistance to the airflow in the workstations. Easy to clean surface. Easy to use

Easy to carry

The aluminum block covers 2/3 of the tube which gives better uniform heat to the tubes.

Sayais dentenje : 21/119

Professor & Head Das & Obts

Professor & Head Das & Obts

Department of Change Likely.