Government of West Bengal Office of the Principal NRS Medical College, Kolkata

Memo No: NMC/Store/ 1033

Date: 28/02/2019

<u>Tender Notice</u> (NIT NO. 79 /2018-19)

Sealed Tender is invited by the Principal, N. R. S. Medical College, Kolkata – 14 for supply of 1) IVF Work Station for the Department of G & O for IVF Centre of this Institution. Tender form and details/specification can be obtained from the Departmental website www.wbhealth.gov.in from 01/03/2019. Last date of submission of Tender is 11/03/2019 up to 1 P.M. Tender will be opened on same date at 2.30 P.M in the office of the Principal, N. R. S. Medical College. Details are available in the Department of Health & Family Welfare, Govt. of West Bengal official Website: www.wbhealth.gov.in & College website (www.nrsmc.edu.in)

Principal
NRS Medical College,
Kolkata-14

Dated: 28/02/2019

Memo No: NMC/Store/ 1033 1 (4)

The Medical Superintendent-cum-Vice Principal, NRSMCH, Kolkata

2. Accounts Officer, NRS Medical College, Kolkata

3. I. T. Cell, Swasthya Bhawan, Salt-lake, for publication in the website.

Notice Board.

Principal
NRS Medical College
Kolkata-14

Terms and Condition

- 1. Only Manufacturers/direct Importers/Authorized Distributors are eligible for quoting.
- should be quoted in both figure and word strictly per sample/mentioned specification along with provided accounting unit only. All rates quoted should be inclusive of all liveable charges (including door delivery charges to the concerned department, charges, installation charges), and all other foreseeable incidental charges are packing exclusive of all applicable taxes should be mentioned separately.
- 3. The Tendering authority reserves the right to withdraw any item from the Tender at any stage. In such Contingency the selection of such item, if already made in favour of any Tendered shall be treated as cancelled. The Tendering authority also reserves the right to accept / reject / split any.
- 4. The Tenderer shall be bound for total replacement of the supplied quantity with a fresh supply of equivalent quantity if the item(s) are found to be defective/not as per given specification.
- Payment will be made only after the delivery of the articles and satisfactory certification by an authority fixed by the administration. No request for advance or part payment will be considered.
- 6. In case of rejection of any supply on quality grounds or failure or refusal to replace the contracted items within the specified time or according to specifications/quality, the College/Hospital administration will have the authority to recover the losses, if any, by making deductions from any pending claim of the supplier in respect of same tender or any past/future tender.
- 7. The Specification and Brochure is to be submitted along with the Tender.
- 8. The Tenderers must accompany all the papers/documents along with the Tender.
- 9. The Tenderers must submit their tender in two bid method. Main envelope should contain Technical bid in Envelop A and Financial bid in Envelop B. all the envelops must be properly sealed before submission.
 - 10. Price bid of those bidders will only be opened who will qualify in technical scrutiny by tender selection committee.
- 11. Earnest money amounting to Rs. 5,000/- (Five thousand) only must be deposited in the form of Demand Draft from any Nationalized Bank in favour of 'Principal, N. R. S. Medical College' payable at Kolkata by L1 bidder before collection order.
- 12. The performance bank guarantee will be mandatory for all suppliers and will not be waived in any case. The successful renderers shall be required to furnish the 'performance bank guarantee' from any Nationalized/scheduled Bank acceptable to Govt. Of West Bengal for one year @ 10% of quoted base rate per item for which the tenderer has been selected as supplier.
- 13. The tendering authority has the liberty to ask for **performance demonstration** of any item during any stage of the tender process, before placing order for supply. In case of dissatisfaction, the authority will have the right to disapprove the offer of L1 bidder and go for the next bidder. Orders for supply of the approved products will be placed with the successful tenderers after execution of agreements and such supply shall have to be made in such installments as may be fixed or spread over the period to be specified in the supply orders to be made in pursuance of the agreements. The equipment is to be transported in such packaging so as to avoid damage to primary package of manufacturer. The successful tenderer will have to supply within the specified time schedule that had been assured at the time of selection as supplier.

14. Guarantee / Warranty Period:

The machine/Instrument should be covered by free on-site comprehensive Warranty for one year after successful installation. The tenderers must quote for 5 (Five) years' comprehensive on-site AMC of entire system (including all spares & labour) after the date of completion of free warranty period.

15. Performance Bank Guarantee:

The performance bank guarantee will be mandatory for all suppliers and will not be waived in any case. The successful tenderers shall be required to furnish the 'Performance Bank Guarantee' from any Nationalized/Scheduled Bank acceptable to Govt of West Bengal for One year @ 10% of quoted base rate per item for which the tenderer has been selected as supplier. The performance bank guarantee will have to be renewed after completion of Free Warranty period, for additional 5 (Five) years under the paid CAMC period before transfer of the CAMC amount to bidder's account.

16. In case of more than one L 1 bidder, weightage will be given to document number 7 of the following list.

List of essential documents to be submitted

- 1. Application submitted in Annexure-I
- 2. Copy of GST Registration certificate showing Registration No. (not provisional)
- 3. Copy of Valid Trade license
- 4. Copy of PAN Card of the bidder
- 5. Copies of Income Tax return for the last three years
- 6. Copies of Profit & Loss A/c and Balance Sheet for the last three years .
- 7. Credential of the bidder in successful supply to Govt. hospitals

Principal
NRS Medical College
Kol-14

ANNEXURE-I

APPLICATION FORMAT

(To be furnished in the Company's official I To The Principal, Nil Ratan Sircar Medical C	letter pad with full address and contact no etc) College, Kolkata-700014.	
	ne of the Item -1)2	
College, Kolkata-700014. Ref:- Sir, Having examined the pre-qualification & or submit all the necessary information and ref. That the application is made by me / us in the capacity of	on behalf ofduly authorized to submit the offer. The attached in Annexure II. aid down in the NIT mentioned above and shout the tender period and its extensions, if any. em /items and assured supply to the Health & of West Bengal as per Table-1 (Instrument ould be raised in our name.	
	(Address, contact no and e mail address) as Others (strike out whichever is not applicable) as	
as prescribed is attached in annexure III 5. In the event of being selected, supply wi excepting the condition which is beyond or 6. We understand that: (a) Tender Selection Committee/ H&FW De under this project.	ur control. ept can amend the scope & value of the contract bid ept reserves the right to reject any application	
Contact details -	Signature of applicant including title and capacity in which application is made.	

E mail ID:

IVF WORKSTATION

- Stainless Steel or Hi-macTable Top, rust free.
- Integrated heat System for table.
- Mounting Assembly for any make Stereozoom Microscopes.
- Microscope Light Opening with Heated Glass.
- Integrated Gassing and Humidification System. Variable FAN speed.
- HEPA Filter alarm indicator. Noise level less than 53dba. Light damping facility to adjust light
- The air flow should filtered through a VOC pre filter and HEPA filter to ensure filtered Grade A air quality.
- HEPA filter classification H-14 according to DIN EN 1822 removes at least 99.995 % of all airborne particles down to 0.3 microns.
- Temperature sensors to ensure high temperature stability at +/- 0.4 oC over the work
- The unique design ensures excellent heat transfer to the samples and avoidance of temperature overshoot.
- The unit temperature is constantly being displayed on the front panel of the work area.
- Build-in Micro CO2 Incubator is essential required to optimize Work conditions inside the Work area, and to reduce the frequency of opening of large CO₂ incubators O2 Incubator can be an Integrated part of the Workstation.
- The clear LED light gives you the highest intensity in the visual field.
- Durable up to 40.000 hours of continuous light. Easily adjustable dark and light field illumination Low power consumption.
- Stainless steel or HI-mac working chamber working table top. The pre-mixed gas is humidified through a heated gas washing system and extruded through nozzles in the table plate covered by glass hoods (glass hoods optional). With the regulator you can adjust the level of humidification by adjusting the gas flow. The gas flow can be adjusted between 0-50 ltr/h.
- Size: 4×2 feet. Preparation for built in Microscope

Department of Cynae. & Obis

Languis Mukuyis.