

GOVERNMENT OF WEST BENGAL  
Office of the Principal  
NRS Medical College, Kolkata-14

E-mail: [principalnrsmc@gmail.com](mailto:principalnrsmc@gmail.com)

phone no : 2286-0033(direct)

Memo No: NMC/Store/ 488

Date: 28/01/2025


Quotation Notice

Sealed quotations are invited from bonafide firms with superscription "Quotation for Article " for the Department of Anatomy of this Institution . The quotation should reach the office of the undersigned on or before 05.02.25 by 1 PM. The quotation should accompany current valid trade Licence, Income Tax Return, PAN No., GST Registration Certificate (not provisional)

Acceptance of the lowest rate is not obligatory for the undersigned that reserves the right to accept or reject any offer without assigning any reason whatsoever. All items should be good quality.

Rate must be submitted after calculating GST or other Taxes, if any according to the following format


Sl. No.	Name Of Articles	RATE/UNIT
1	Advanced rotary microtome	

  
Principal  
NRS Medical College, Kolkata-14

Memo No: NMC/Store / 488/1 (A)

Dated: 28/01/2025

1. The Dy. Director of Health Services(E&S),Govt of West Bengal, CMS, Kolkata
2. The Joint Secretary, IT Cell, Department of Health and Family Welfare. Govt. Of West Bengal, Swasthya Bhawan, Kol-14, with the request for publication of the Quotation Notice and Terms & Condition in the Govt Website.
- 3.-5. The Principal, Medical College/R.G.Kar Medical College/National Medical College, Kolkata
6. The District Magistrate, South 24 pargana, Howrah, Nadia, West Bengal.
7. The Medical Superintendent-cum-Vice Principal, NRSMCH, Kolkata
8. The Director ,IPGME&R, Kolkata
9. Notice Board

  
Principal  
NRS Medical College, Kolkata-14



### Specification of Advanced Rotary Microtome

Operation.	Manual.
Section Thickness Range.	1 $\mu$ m To 60 $\mu$ m.
Increment Value.	In Increment Of 1 $\mu$ m.
Specimen Orientation.	8 <sup>o</sup> X,y Axis, 360 <sup>o</sup> Around Z Axis.
Vertical Stroke.	70mm.
Horizontal Advancement.	25mm.
Maximum Sectioning Area.	50mm x 50mm.
Dimension (L X W X H)	480mm x 390mm x 280mm.
Weight.	34 kg (approx)
Two Step Trimming.	Optional 15 $\mu$ m, 25 $\mu$ m.
Retraction.	Optional.
Universal Cassettes Clamp.	Optional.

*Shajumdas* 04.02.2025.

Professor & Head  
Department of Anatomy  
N.R.S. Medical College  
Kolkata-700014



- Counter Weight Balanced Hand Wheel.
- Ergonomically Design Course Feed Wheel.
- Ergonomically Located Hand Wheel Lock.
- Lateral Displacement Of Blade Holder To Ensure The Use Of Entire Length Of Blade.
- Acoustic And Acroscopic Signal Of Specimen Feed At End And Home Position.
- Low And High Profile Compatibles Blade Holder.
- Blade Holder With Integrated Finger Protect Guard And Blade Ejector.
- Spacious Waste Tray.

*Majumdar* 04.02.2025

Professor & Head  
Department of Anatomy  
N.R.S. Medical College  
Kolkata-700

*Majumdar* 04.02.2025

Professor & Head  
Department of Anatomy  
N.R.S. Medical College  
Kolkata-700014