

Government of West Bengal
Office of the Principal
NRS Medical College, Kolkata

Memo No: NMC/Store/ 2629

Date: 27/5/19

Tender Notice
(NIT NO. 11 /2019-20)

Sealed Tender is invited by the Principal, N. R. S. Medical College, Kolkata – 14 for supply of **1) High flow nasal cannula** for the Department of Pediatric of this Institution. Tender form and details/specification can be obtained from the Departmental website www.wbhealth.gov.in from 28/05/2019. Last date of submission of Tender is 10/06/2019 up to 1 P.M. Tender will be opened on same date at 2.30 P.M in the office of the Principal, N. R. S. Medical College. Details are available in the Department of Health & Family Welfare, Govt. of West Bengal official Website: www.wbhealth.gov.in & College website (www.nrsmc.edu.in)


Terms and Condition

1. Only Manufacturers/direct Importers/Authorized Distributors are eligible for quoting.
2. The rate should be quoted in both figure and word strictly as per sample/mentioned specification along with provided accounting unit only. All rates quoted should be inclusive of all liveable charges (including door delivery charges to the concerned department, packing charges, installation charges), and all other foreseeable incidental charges are exclusive of all applicable taxes should be mentioned separately.
3. The Tendering authority reserves the right to withdraw any item from the Tender at any stage. In such Contingency the selection of such item, if already made in favour of any Tendered shall be treated as cancelled. The Tendering authority also reserves the right to accept / reject / split any.
4. The Tenderer shall be bound for total replacement of the supplied quantity with a fresh supply of equivalent quantity if the item(s) are found to be defective/not as per given specification.
5. Payment will be made only after the delivery of the articles and satisfactory certification by an authority fixed by the administration. No request for advance or part payment will be considered.

6. In case of rejection of any supply on quality grounds or failure or refusal to replace the contracted items within the specified time or according to specifications/quality, the College/Hospital administration will have the authority to recover the losses, if any, by making deductions from any pending claim of the supplier in respect of same tender or any past/future tender.
7. The Specification and Brochure is to be submitted along with the Tender.
8. The Tenderers must accompany all the papers/documents along with the Tender.
9. The Tenderers must submit their tender in two bid method. Main envelope should contain Technical bid in Envelop A and Financial bid in Envelop B. all the envelopes must be properly sealed before submission.
10. Price bid of those bidders will only be opened who will qualify in technical scrutiny by tender selection committee.
11. Earnest money amounting to Rs. 5,000/- (Five thousand) only must be deposited in the form of Demand Draft from any Nationalized Bank in favour of 'Principal, N. R. S. Medical College' payable at Kolkata by L1 bidder before collection order.
12. The performance bank guarantee will be mandatory for all suppliers and will not be waived in any case. The successful tenderers shall be required to furnish the 'performance bank guarantee' from any Nationalized/scheduled Bank acceptable to Govt. Of West Bengal for one year @ 10% of quoted base rate per item for which the tenderer has been selected as supplier.
13. The tendering authority has the liberty to ask for **performance demonstration** of any item during any stage of the tender process, before placing order for supply. In case of dissatisfaction, the authority will have the right to disapprove the offer of L1 bidder and go for the next bidder. Orders for supply of the approved products will be placed with the successful tenderers after execution of agreements and such supply shall have to be made in such installments as may be fixed or spread over the period to be specified in the supply orders to be made in pursuance of the agreements. The equipment is to be transported in such packaging so as to avoid damage to primary package of manufacturer. The successful tenderer will have to supply within the specified time schedule that had been assured at the time of selection as supplier.
14. **Guarantee / Warranty Period:**
The machine/Instrument should be covered by free on-site comprehensive Warranty for one year after successful installation. The tenderers must quote for 5 (Five) years' comprehensive on-site AMC of entire system (including all spares & labour) after the date of completion of free warranty period.
15. **Performance Bank Guarantee:**
The performance bank guarantee will be mandatory for all suppliers and will not be waived in any case. The successful tenderers shall be required to furnish the 'Performance Bank Guarantee' from any Nationalized/Scheduled Bank acceptable to Govt of West Bengal for One year @ 10% of quoted base rate per item for which the tenderer has been selected as supplier. The performance bank guarantee will have to be renewed after completion of Free Warranty period, for additional 5 (Five) years under the paid CAMC period before transfer of the CAMC amount to bidder's account.
16. Existing financial norms will be followed for accepting / cancellation of any bid.
17. In case of more than one L 1 bidder, weightage will be given to document number 7 of the following list.

List of essential documents to be submitted

1. Application submitted in Annexure- I
2. Copy of GST Registration certificate showing Registration No. (not provisional)
3. Copy of Valid Trade license
4. Copy of PAN Card of the bidder
5. Copies of Income Tax return for the last three years
6. Copies of Profit & Loss A/c and Balance Sheet for the last three years .
7. Credential of the bidder in successful supply to Govt. hospitals



Principal
NRS Medical College
Kol-14

ANNEXURE-I

APPLICATION FORMAT

(To be furnished in the Company's official letter pad with full address and contact no etc)
To The Principal, Nil Ratan Sircar Medical College, Kolkata-700014.

Sub: Tender No.-----Dt.------(Name of the Item -1)-----2)-----
-----3)-----4)-----

----) for purchase and installation of equipment for department of, Nil Ratan Sircar Medical College, Kolkata-700014.

Ref:-

Sir,

Having examined the pre-qualification & other documents published in the N.I.T, I /we hereby submit all the necessary information and relevant documents for evaluation:

1. That the application is made by me / us on behalf of.....

In the capacity of.....duly authorized to submit the offer. The authorization letter from the Company is attached in Annexure II.

2. We accept the terms and conditions as laid down in the NIT mentioned above and declare that we shall abide by it for throughout the tender period and its extensions, if any.

3. We are offering rate for the following item /items and assured supply to the Health & Family Welfare Department, Government of West Bengal as per Table-1 (Instrument Specification)

4. a. We propose that the order and bill should be raised in our name.

For liasoing, we have appointed M/S having its office at (Address, contact no and e mail address) as C&F agent /Liasoner /Contact person/..... Others (strike out whichever is not applicable) as per clause of the NIT

OR

c. The agreement between ourselves and the distributor/C&F/Liasoner& other documents as prescribed is attached in annexure III

5. In the event of being selected, supply will be made within the stipulated period excepting the condition which is beyond our control.

6. We understand that:

(a) Tender Selection Committee/ H&FW Dept can amend the scope & value of the contract bid under this project.

(b) Tender Selection Committee/ H&FW Dept reserves the right to reject any application without assigning any reason.

Date :-

Signature of applicant including title
and capacity in which application is made.

Contact details -


Mobile:

E mail ID:

Specification of High flow nasal cannula

Technical Specification – Heated Humidified High Flow Nasal Cannula system – In Adult, Infant & Paediatric

1. The device should have integrated flow generator to deliver flows from 2 to 25 ltrs in Pediatric and 10 to 60 ltrs in Adult settings.
2. Integrated Air/oxygen blending and Fio2 monitoring with facility to deliver wide range of oxygen concentrations from 21 to 100 %.
3. Oxygen sensor does not require in field calibration.
4. Clinician menu for setting range of Fio2 and flow settings.
5. Inbuilt heated humidifier to deliver warm and humid gases to airway. The humidification chamber should have dual float with auto feed system.
6. Display to monitor temperature of humidified gas, flow rate and fio2
7. Visual and audible alarm indication for :
 - a. Tubes disconnect Leaks, tube blockages, and water out and hardware fault with error codes. Audible power failure alarm.
8. The device should have thermal disinfection mode to minimize contamination. Heated tube for sterilization of the device should be provided with the device.
9. Accessories to be supplied with the system:
 - Patient breathing tube with integrated heating coil and flow/temperature sensor.
 - Soft, flexible anatomically contoured nasal cannula. The cannula should be available in different sizes suitable for infant and paediatric application. It should have hydrocolloid adhesive on Velcro to secure on skin. This should be easily replaceable.
 - Air filter – 2 nos
 - Disinfection kit – 1no
 - Oxygen inlet extension tube- 1no
 - Mounting tray and pole with castors & IV hook
- 9) Supply frequency: 50/60 Hz, Supply voltage: 220-240 v
- 10) Sound pressure level 45dbA@1m
- 11) Oxygen analyzer accuracy +/-10 % (within range of 25-80 % Fio2)
- The unit should be compliance with international standards. IEC 60601 and ISO 13485 quality standard
10. Certification: US FDA and European CE approved
11. Compliant with international standards. IEC 60601 and ISO 13485 quality
12. Suitable for use in ICU, HDU, recovery, wards and emergency department.


PROFESSOR AND HOD
Department of Pediatrics
Niralan Sirsar Medical College & Hospital
Kolkata-14